

### Tissue Glue

Catalog Number: ACR-022

Document #: ACR-022

Effective Date: 12/04/2017, IFU-151 Rev A

**Intended Use** 

For In Vitro Diagnostic Use.

#### **Product Description**

Tissue Blue reacts with silica glass to change its surface chemistry so that is may bond either ionically or covalently with aldehyde and ketone groups in tissue sections.

### **Summary and Explanation**

Proper tissue adhesion to glass slides is a critical step in immunohistochemistry. Good immunostaining may require harsh pretreatment of tissues for antigen retrieval (e.g. proteolytic enzyme digestion, boiling in high pH buffer, treatment with detergents) Improper adhesion of tissue sections can lead to poor quality staining or even loss of tissue sections. White glue, albumen, chrome-gelatin, and poly-L-lysine increase background staining. There is a need for a tissue adhesive that can irreversible bind tissue sections to glass without creating background staining. Glass treated with Tissue Glue can be used for both paraffin and frozen tissue sections.

#### Format

Slightly viscous, ready to use clear solution

## Volume/UOM

7 mL

# **Storage and Handling**

Store at 2-8°C. Do not use after expiration date printed on label.

### **Preparation of Working Solutions**

- Work in a chemical fume hood with appropriate ventilation and personal protective equipment.
- 2. Once the bottle has been opened, use it immediately.
- Prolonged storage is not recommended. Discard unused working solution
- 4. Working solution should be prepared.

#### **Protocol Recommendations**

- Place slides in a metallic or acetone-resistant slide holder. Clean slides by immersing them in acetone for 1 minute.
- 2. Immerse slides in Tissue Glue working solution for 30 seconds.
- Remove and drain slides. Rinse slides thoroughly with deionized water three times.
- 4. Change the water after every rinse.
- 5. Avoid creating water bubbles by gently agitating the slide holder.
- 6. Air dry slides at room temperature
- Once dried, slides can be used immediately or can be stored in a box at room temperature.

### **Quality Control**

Refer to CLSI Quality Standards for Design and Implementation of Immunohistochemistry Assays; Approved Guideline-Second edition (I/LA28-A2). CLSI Wayne, PA, USA (www.clsi.org). 2011

#### Troubleshooting

Contact StatLab Medical Products Technical Support at (800) 442-3573, option 5 or ihctech@statlab.com to report unusual staining.

#### Warranty

There are no warranties, expressed or implied, which extend beyond this description. StatLab Medical Products is not liable for property damage, personal injury, or economic loss caused by this product.

#### **Performance Characteristics**

The protocols for a specific application can vary. These include, but are not limited to: fixation, heat-retrieval method, incubation times, tissue section thickness and detection kit used. Due to the superior sensitivity of these unique reagents, the recommended incubation times and titers listed are not applicable to other detection systems, as results may vary. The data sheet recommendations and protocols are based on exclusive use of StatLab Medical products. Ultimately, it is the responsibility of the investigator to determine optimal conditions. These products are tools that can be used for interpretation of morphological findings in conjunction with other diagnostic tests and pertinent clinical data by a qualified pathologist.

## **Precautions**

- 1. Wear disposable gloves when handling reagents.
- Specimens, before and after fixation, and all materials exposed to them should be handled as if capable of transmitting infection and disposed of with proper precautions. Never pipette reagents by mouth and avoid contacting the skin and mucous membranes with reagents and specimens. If reagents or specimens come in contact with sensitive areas, wash with copious amounts of water.
- Microbial contamination of reagents may result in an increase in nonspecific staining.
- 4. Incubation times or temperatures other than those specified may give erroneous results. The user must validate any such change.
- 5. Do not use reagent after the expiration date printed on the label.
- 6. The SDS is available upon request.
- Consult OSHA, federal, state or local regulations for disposal of any toxic substances.

